

Urgent Recall Notification
Please Read Immediately

September 17, 2020

Dear Valued Customer;

Perrigo is conducting a retail level recall of all unexpired batches of Albuterol Sulfate Inhalation Aerosol. This recall is being conducted due to complaints indicating that some units may not dispense due to clogging.

This voluntary action is being taken with the knowledge of the FDA and is being issued to the retail level. We are asking for returns of the affected batches from all warehouse and retail/pharmacy locations.

In accordance with FDA guidelines, Perrigo must verify notifications in order to determine the effectiveness of this recall. Perrigo will begin making these Recall Effectiveness Checks by phone within three to four weeks.

We deeply regret any inconvenience and appreciate your assistance with this recall. Please contact our Perrigo Recall Information Line at 877-907-9964 if we can provide additional information or assistance.

Sincerely,

Handwritten signature of David Tyree, consisting of a stylized 'D' and 'T' followed by a flourish.

David Tyree
Vice President of Rx Quality Operations

Attachment: Recall Notification

Copy: Jim Dillard III, Chief Scientific Officer
Sharon Kochan, Executive Vice President and President Rx Business
Ron Hubbert, Recall Coordinator

Recall Notification
Albuterol Sulfate Inhalation Aerosol
(Retail Level)
September 2020

Products	<p>Product(s): Albuterol Sulfate Inhalation Aerosol</p> <p>NDC#: 45802-088-01</p> <p>UPC#: 345802088017</p> <p>Distribution Dates: 02/26/20 to 08/21/20</p> <p>See attached List of Affected Batches</p>
Reason	<p>These products are being recalled due to complaints indicating that some units may not dispense due to clogging.</p>
Action	<ol style="list-style-type: none">1. Immediately examine your inventory and quarantine product subject to recall. Check your warehouses and retail locations and remove affected product from distribution.2. Forward this recall notification to any retail locations that may have received the affected product.3. Complete the Business Reply Card and return as directed on the form.4. Effectiveness Checks: The FDA requires Perrigo to verify notifications to determine the effectiveness of this recall. Perrigo will begin making these Effectiveness Checks by phone within three to four weeks. In order to complete this step, Perrigo requires a list of the warehouse and retail locations that may have received the affected products. Please return the list of warehouse and retail locations by fax to 269-686-3751.5. Return the product within 30 days to: Stericycle Pharmaceutical Services 2670 Executive Drive Suite A Indianapolis, IN 46241 Contact Stericycle for directions on returning the affected product: Event #: 3279 IVR: 877-907-9964 FAX: 877-523-9109 Email: perrigo3279@stericycle.com6. Individual Returns: Package your return in a cardboard box, return must not exceed 1 Liter. Once packaged must attach the enclosed Limited Quantity label to the outside of the package before shipping. Pallet Returns: Must not exceed 550lbs. Must attach the Limited Quantity label to each pallet and shrink wrap.7. Include RA number 60346340 with all returns.

Credit	Credit will be issued for product returned to Stericycle under the correct RA # and matching the product's description above. Credit will be provided when customers request a deduction. Credit will only be issued for product from the affected lots returned within 30 days.
Other Information	<p>This recall is only for the product listed on the attached report.</p> <p>Notifications are being sent out to all the accounts that purchased this product from Perrigo.</p> <p>This recall is to the retail/pharmacy level only. Any patient who wishes to return their prescription will be asked to speak with their pharmacist about exchanging product.</p> <p>We sincerely regret any inconvenience caused by this action and appreciate your immediate attention and cooperation.</p> <p>For assistance or questions about this Recall process please contact the Perrigo Recall Information Line at 877-907-9964.</p>

Albuterol Recall Batch List

Batch	Expiration
18MC-052	09/2020
18MC-055	09/2020
18MC-056	09/2020
18MC-057	09/2020
18MC-058	09/2020
18MC-060	09/2020
18MC-061	10/2020
18MC-062	10/2020
18MC-064	10/2020
18MC-065	10/2020
18MC-066	10/2020
18MC-068	10/2020
18MC-069	10/2020
18MC-070	10/2020
18MC-071	10/2020
18MC-072	10/2020
18MC-073	10/2020
18MC-074	10/2020
18MC-075	11/2020
18MC-076	11/2020
18MC-077	11/2020
18MC-078	11/2020
18MC-079	11/2020
18MC-080	11/2020
18MC-081	11/2020
18MC-082	11/2020
18MC-083	11/2020
18MC-084	11/2020
18MC-085	11/2020
18MC-095	12/2020
18MC-096	12/2020
18MC-097	12/2020
18MC-098	12/2020
18MC-099	12/2020
19MC-001	1/2021
19MC-002	1/2021
19MC-003	1/2021
19MC-004	1/2021
19MC-005	2/2021

19MC-006	2/2021
19MC-007	2/2021
19MC-008	2/2021
19MC-009	2/2021
19MC-010	2/2021
19MC-011	2/2021
19MC-012	2/2021
19MC-013	2/2021
19MC-014	2/2021
19MC-015	2/2021
19MC-020	4/2021
19MC-041	4/2021
19MC-042	4/2021
19MC-043	4/2021
19MC-044	4/2021
19MC-045	4/2021
19MC-046	5/2021
19MC-047	5/2021
19MC-048	5/2021
19MC-049	5/2021
19MC-050	5/2021
19MC-051	5/2021
19MC-052	5/2021
19MC-053	6/2021
19MC-054	6/2021
19MC-055	6/2021
19MC-056	6/2021
19MC-057	6/2021
19MC-059	6/2021
19MC-060	6/2021
19MC-061	7/2021
19MC-062	7/2021
19MC-063	7/2021
19MC-064	7/2021
19MC-065	7/2021
19MC-066	7/2021
19MC-067	7/2021
19MC-068	7/2021
19MC-069	7/2021
19MC-070	7/2021
19MC-083	7/2021
19MC-084	7/2021
19MC-085	7/2021
19MC-087	8/2021
19MC-088	8/2021

19MC-090	8/2021
19MC-091	8/2021
19MC-092	8/2021
19MC-093	8/2021
19MC-094	8/2021
19MC-095	8/2021
19MC-096	8/2021
19MC-097	8/2021
19MC-098	8/2021
19MC-099	8/2021
19MC-101	9/2021
19MC-102	9/2021
19MC-103	9/2021
19MC-104	9/2021
19MC-107	9/2021
19MC-108	9/2021
19MC-109	9/2021
19MC-110	9/2021
19MC-112	9/2021
19MC-113	9/2021
20MC-014	3/2022
20MC-015	3/2022
20MC-016	3/2022
20MC-017	3/2022
20MC-020	3/2022
20MC-021	3/2022
20MC-022	4/2022
20MC-023	4/2022
20MC-024	4/2022
20MC-029	4/2022
20MC-030	4/2022
20MC-032	4/2022
20MC-034	4/2022
20MC-035	4/2022
20MC-036	4/2022
20MC-037	4/2022
20MC-038	4/2022
20MC-039	4/2022
20MC-040	4/2022
20MC-041	4/2022
20MC-042	4/2022
20MC-043	4/2022
20MC-044	4/2022
20MC-045	5/2022
20MC-046	5/2022

20MC-047	5/2022
20MC-048	5/2022
20MC-049	5/2022
20MC-050	5/2022
20MC-051	5/2022
20MC-052	5/2022
20MC-053	5/2022
20MC-054	5/2022
20MC-061	5/2022
20MC-062	5/2022
20MC-063	5/2022
20MC-064	5/2022
20MC-065	5/2022
20MC-066	5/2022
20MC-067	5/2022
20MC-068	5/2022
20MC-069	6/2022
20MC-070	6/2022
20MC-072	3/2022
20MC-078	6/2022
20MC-079	6/2022
20MC-080	6/2022
20MC-081	6/2022
20MC-083	6/2022
20MC-084	6/2022